

# LVS-7500 SERIES

## 21 CFR Part 11 Compliance Explained

*CFR is an abbreviation for Code of Federal Regulations and is used by the USA Government to define legal regulations. 21 CFR is the general classification within the regulation. Part 11 is the part that defines what we are concerned about in this 21 CFR. In this article, Part 11 is used to mean the entire 21 CFR Part 11; it defines the record keeping of testing as it relates to the process of making drugs, medical equipment, etc. 21 CFR Part 11 covers the paperwork trail necessary in tracking the production life cycle of items being manufactured, including the signatures of the Quality people indicating that a process*

*was completed and tested, and either meets the standards or did not meet standards. In the past, these records were on paper, which became the permanent part of the records required by the Food and Drug Administration (FDA) should they need to examine the production process records for any reason.*

*Now days, record processing is performed electronically so there is no paper trail for the FDA to examine. Part 11 addresses this item as far as providing an electronic signature of the test performed.*

### **LVS-7500 and 21 CFR Part 11**

On the LVS-7500, user names and passwords must be entered, which constitutes an electronic signature. There are a number of requirements the LVS-7500 also meets, such as keeping track of detail records, audit trails of the changes, and more, all of which are required by Part 11. Microscan states that our systems are Compliant READY. We do not say they are COMPLIANT; there is a big difference between the two statements. If an item is COMPLIANT, it must be registered with the FDA and meet extensive requirements that are not the remit of the LVS-7500 product. The LVS-7500 is a Commercial Off-The-Shelf product (COTS), meaning it is not customized to any system; it is as it is. Buyers of the system are responsible for the integration and validation of the LVS-7500 product into their compliant system.

### **National Guidelines**

The buying organization of the system with FDA compliance is responsible to the FDA for validation; they need to have in place the procedural and administrative controls as required by cGMP, ISPE GAMP<sup>®</sup>, FDA or other National guidelines. The LVS-7500 has the required abilities to meet the requirements to make the system compliant by the purchasing organization. This includes, but is not limited to the following:

- Ensures accuracy, reliability and consistent intended performance
- The ability to generate accurate, complete copies of records in human readable and electronic form suitable for inspection, review, and copying
- Limiting system access to authorized individuals
- Use of secure, computer-generated, time-stamped

audit trails to independently record the date and time of operator entries and actions that create, modify, or delete electronic records

- Use of authority checks to ensure that only authorized individuals can use the system, electronically sign a record, or perform operations at hand
- Record changes shall not obscure previous recorded data

• And much more!

LVS-has available for separate purchase a Validation Procedure Outline that covers Installation Qualification (IQ) and Operational Qualification (OQ) protocol. The outline is available for purchase and use by the buying organization to customize for their validations protocol.

**SAFETY CERTIFICATIONS DESIGNED FOR**  
FCC, CE, UL (Pending)

**ROHS COMPLIANT**

**QMS CERTIFICATION**

[www.microscan.com/quality](http://www.microscan.com/quality)

©2016 Microscan Systems, Inc. ML050A-EN-0416

Warranty – For current warranty information about this product, please visit [www.microscan.com/warranty](http://www.microscan.com/warranty).



**MICROSCAN**  
[www.microscan.com](http://www.microscan.com)